



## Instruction for Use – Spray Catheter for Wide Spray SU

### Attention:

**This medical device may only be purchased by specialists, doctors and medical personnel and may only be used in accordance with these instructions for use and the defined area of application.**

### Description / Versions

ENDO-FLEX GmbH offers different Spray Catheters for Wide Spray in different lengths for various application cases: 120 / 180 / 230 cm. Catheter diameter is 1.8 mm for the lengths.

### Products

This user manual is valid for the products listed below:

- E47185-A
- E47185-C
- E47185-G

### Important Note

Please ensure you read this user manual carefully before each use and keep it somewhere that is easily accessible for users and/or appropriate specialist staff.

**Ensure you carefully read through the warnings. Patients, users or third parties may suffer serious injury as a result of the improper use of the products.**

### Content and Packaging

- 1 outer box
- 5 Spray Catheter SU (Single Use) sterile packed
- 1 Instructions for use

### Patient Population

The patient population or target patient group is derived from the indication determined by the responsible doctor who diagnostically or therapeutically treats the patient as part of an endoscopic procedure (the leading procedure per se) according to the intended use of the medical device.

There are no known restrictions to the patient population or the target patient group.

#### • Use of the product on minors:

The user can use the product on minors in the case that the physiological and anatomical conditions of the patient permit the use of the product.

#### • Use of the product on women who are pregnant or breastfeeding:

The indication for the use of the product on women who are pregnant or breastfeeding must be narrowed down by the user based on the respective individual physiological and anatomical conditions of the patient.

### 1. Scope of Application

The products listed above must only be used by physicians with appropriate specialist training in gastroenterology. The products are intended solely for the medical sector as illustrated below and must therefore be used within an operating environment suitable for this purpose. It is essential that both the user as well as the appropriate specialist staff familiarize themselves with the instruments before the user makes use of them.

### 2. Application Period

Spray Catheters are intended to be used uninterrupted for a period of less than 60 minutes under normal conditions (MDD 93/42 EEC).

### 3. Intended Use

Spray Catheters for Wide Spray are used for applying anaesthetics during a bronchoscopy and for dying suspicious mucosal lesions which are very difficult to evaluate during routine endoscopy by spraying colour. The use is carried out by means of an endoscope.

#### Indications

- Endoscopic intravital staining
- applying anaesthetics

#### Contraindications

- Not fasting patients
- Fragility of the intestinal wall: e.g. highly florid inflammation of the colon (e.g. ulcerative colitis, diverticulitis, ulcerative colitis, toxic megacolon)
- Peritonitis, acute abdomen, e.g. intestinal perforation, ileus
- Sepsis
- Co-morbidity, e.g. severe cardiopulmonary diseases and decompensation
- Uncontrollable haemorrhagic diatheses
- Pregnancy
- Recently created gastrointestinal anastomosis.

### 4. Complications / Side Effects / Cross-reactions

Sedation during endoscopic examination increases the risk of hypoxemia, hypercapnia, hypotension, arrhythmias and aspiration due to reduced protective reflexes. Hypoxemia also occurs without sedation during endoscopic examinations due to the feed of the endoscope. Possible injuries in connection with endoscopic examinations can be:

- Perforation
- Hemorrhage

- Fever
- Infection
- Allergic reaction to medication
- Hypertension
- Respiratory depression or arrest
- Cardiac arrhythmia or arrest
- Hypoventilation
- Bronchospasm
- Laryngospasm
- Hypoxemia
- Barotraumas

Corresponding preparations for occurring complications are to be done before starting the application!

### 5. Warnings / Precautions

These instructions must be followed, as well as instructions from compatible components and hospital regulations for infection prevention, safe use, cleaning and sterilization. Failure to do so may result in serious injury to the patient and/or the user.

The following applies to the product:

- All components should be carefully checked for compatibility and integrity before use. Do not use defective instruments! If defects occur, dispose of the instrument and replace it with a new one.
- Never use the product outside the recommended technical specifications (intended use).
- Never tamper with the structural conditions of the instrument, avoid kinks and other damage, and immediately discontinue use in the event of a malfunction!
- Wear protective clothing (gloves, goggles, gown, etc.) is absolutely necessary!
- Comparison of the technical data of the product with those of the endoscope used. The working channel diameter must be at least 0.2 mm larger than the outer diameter of the instrument.
- Never force instruments into the working channel!

### 6. Compatibility

Not specified.

### 7. Function Test

The medical devices must be checked with regard to the following aspects prior to use:

- Expiration date
- Undamaged packaging
- Damage on the product (cracks, bending, deformation)

Inspect products for immaculate surfaces, correct assembly and functionality.

**Products that have failed the function test must not be used as their sterility and product safety cannot be ensured. Dispose of these products accordingly, or please return them to the manufacturer.**

### 8. Preparation / Application

#### Preparation

- Check the instrument for reliable function and irregularities before the procedure.
- **! If you notice any irregularities, replace the instrument with a new one.**
- The Spray Catheter must be passed through the endoscope operating channel with its mandrin inserted. Please observe compatibility between the catheter and the endoscope operating channel diameter.

#### Application

##### Applying of anaesthetics

- Please observe that the mandrin is firmly fastened to the Luer-Lock adapter.
- Connect a syringe filled with the anaesthetic to the Luer-Lock adapter at the Spray Catheter.
- Insert the Spray Catheter into the operating channel of the broncho-scope and spray the anaesthetic while pushing the plunger of the syringe with constant pressure.

##### Chromoendoscopy

- Please observe that the mandrin is firmly fastened to the Luer-Lock adapter.
- Connect a syringe filled with the chosen dying colour to the Luer-Lock adapter at the Spray Catheter.
- Slowly move the catheter's spray head to the suspicious tissue areas while pushing the plunger of the syringe with constant pressure.
- After the application please wait for approx. 2 min for best results.
- Finally, rinse the treated tissue area with tap water for best visibility of the dyed lesions.

**Note: For best spraying results please use a 10 ml syringe!**

#### Recommended Dying Colours

- Lugolic solution 1 %
- Methylenblue 0.5 – 1 %
- Indigocarmin 0.1 %





GA-0186 - Vers. 6.0 (2020-01)



CE 0483


**ENDO-FLEX®**  
 A MEDI-GLOBE COMPANY

## ***Instruction for Use – Spray Catheter for Wide Spray SU***

### **9. Reprocessing and Sterilisation**

These instruments are delivered in sterile condition and **CANNOT EFFECTIVELY** be cleaned, disinfected and sterilised after single use on account of the design which can no longer be removed and must be disposed of after single use.

The products are only sterile if the packaging is undamaged and unopened and if the shelf life is not exceeded. Products whose packaging is damaged or whose shelf life has expired must be discarded.

### **10. Service Life of Products**

The service life of the product is typically 3 years after the date of manufacture under normal conditions.

### **11. Liability and Warranty**

As the manufacturer here, ENDO-FLEX shall not be liable for any damage or consequential damage arising from improper use or handling. This shall in particular apply to any use not in line with the defined purpose or failure to observe the preparation and sterilisation instructions and warnings. This shall also apply to repairs or modifications to the product made by individuals who are not authorised to do so by the manufacturer.

### **12. Service**

Do not carry out any modifications to the product. If you have any concerns, complaints or comments regarding our products, please get in touch with us.

### **13. Return transport**

Defective or non-conforming products must have gone through the entire reprocessing procedure prior to be sent back for repair/servicing. Please ensure you label the products accordingly with the note "hygienically safe" or "not decontaminated".

### **14. Transport and Storage Conditions**

- Products may only be transported and stored in the packaging provided for this purpose.
- Products must be stored dry and protected from sunlight at room temperature.
- Do not place any objects on the storage packaging and the sterile barrier system!
- Ensure the instruments are not kept near to chemicals, disinfectants or radioactive radiation!

### **15. Disposal**

The products, packaging materials and accessories must all be disposed of in accordance with the nationally applicable regulations and laws. No specific instructions regarding this are given by the manufacturer.

### **16. Symbols used**



Symbol for "Item Number"



Symbol for "Batch code"



Symbol for "Manufacturer"



Symbol for "Date of manufacture"



Symbol for "Observe instructions for use"



Symbol for "Sterilized with ethylene oxide"



Symbol for "non-reuse"



Symbol for "Do not re-sterilize"



Symbol for "Do not use if packaging is damaged"



Symbol for "Expiry Date"



Symbol for "keep dry"



Symbol for "protect from sunlight"



Symbol for "Caution"

